

AQ-SPEC

Air Quality Sensor Performance Evaluation Center

Evaluation Summary

Sensor Description

Manufacturer/Model:
Vaisala/AQT530

Pollutants:
O₃

Time Resolution:
1-min

Type: Electrochemical



- Overall, the accuracy of the Vaisala AQT530 sensors ranged from 64.9% to 94.6% and decreased as O₃ conc. increased over the tested concentration range, except at the first steady state. Overall, the sensors overestimated the O₃ measurements from FEM T400 in the laboratory experiments at 20°C and 40% RH.
- The Vaisala AQT530 sensors exhibited high precision for all T/RH combinations and all O₃ concentrations.
- The Vaisala AQT530 sensors (IDs: 673, 885, 847) showed low to high intra-model variability in the field and laboratory evaluations.
- Data recovery was ~75% - 89% from all units in both field and laboratory evaluations.
- The Vaisala AQT530 sensors showed very weak to weak correlations ($0.22 < R^2 < 0.47$, 5-min mean) with the corresponding FEM T400 data in the field evaluation and very strong correlations with the FEM T400 in the laboratory evaluations ($R^2 > 0.96$).
- The same three Vaisala AQT530 units were tested both in the field (1st stage of testing) and in the laboratory (2nd stage of testing).

Field Evaluation Highlights

- Deployment period 01/14/2022 to 03/25/2022 : the three Vaisala AQT530 sensors showed very weak to weak correlations with the corresponding FEM O₃ data.
- The units exhibited low intra-model variability and data recovery for O₃ measurements was ~86-89% from all units.

Additional Information

Field evaluation report:



<http://www.aqmd.gov/aq-spec/evaluations/criteria-pollutants/field>

Lab evaluation report:

<http://www.aqmd.gov/aq-spec/evaluations/criteria-pollutants/laboratory>

AQ-SPEC website:

<http://www.aqmd.gov/aq-spec>



Coefficient of Determination (R^2) quantifies how the three sensors followed the O₃ concentration change by the reference instruments.

An R^2 approaching the value of 1 reflects a near perfect agreement, whereas a value of 0 indicates a complete lack of correlation.

Laboratory Evaluation Highlights

Accuracy (O₃)

$$A (\%) = 100 - \frac{|\bar{X} - \bar{R}|}{\bar{R}} * 100$$


Steady State (#)	Sensor Mean (ppb)	FEM T400 (ppb)	Accuracy (%)
1	18.5	28.5	64.9
2	50.1	47.5	94.6
3	98.7	88.6	88.6
4	191.4	150.6	72.9
5	331.2	257.0	71.1

Accuracy was evaluated by a concentration ramping experiment at 20°C and 40% RH. The sensor's readings at each ramping steady state are compared to the reference instrument.

A negative % means sensors' overestimation by more than two fold. The higher the positive value (close to 100%), the higher the sensor's accuracy.




Precision (O₃)



100% represents high precision.

Sensor's ability to generate precise measurements of O₃ concentration at low, medium, and high pollutant levels were evaluated under 9 combinations of T and RH, including extreme weather conditions like cold and dry (5°C and 15% RH) cold and humid (5°C and 65% RH), hot and humid (35°C and 65% RH), or hot and dry (35°C and 15% RH).

Coefficient of Determination



The Vaisala AQT530 sensors showed very strong correlations with the corresponding FEM T400 O₃ data ($R^2 > 0.96$) at 20°C and 40% RH.

Climate Susceptibility

From the laboratory studies, temperature and relative humidity had minimal effect on the precision of the Vaisala AQT530 sensors' ozone measurements.

Observed Interferents

NO₂



All documents, reports, data, and other information provided in this document are for informational use only. Mention of trade names or commercial products does not constitute endorsement or recommendation. As a Government Agency, the South Coast AQMD and its AQ-SPEC program highly recommend interested entities to make use and purchase decisions based on the requirements of their study design, the technical aspects and features of their specific project applications.